

USTAR WHOLE-PROCESS SOLUTIONS FOR POINT-OF-CARE DIAGNOSTIC TESTING



Ustar Biotechnologies (Hangzhou) Ltd.



CONTENTS

01



ABOUT USTAR

02



PortNAT® SYSTEM

At-Home
Self-Testing

03



EasyNAT® SYSTEM

Fast Screening and
Primary Diagnosis

04



MultNAT® SYSTEM

Therapy-Oriented
Diagnosis



CONTENTS

01



ABOUT USTAR

02



PortNAT® SYSTEM

At-Home
Self-Testing

03



EasyNAT® SYSTEM

Fast Screening and
Primary Diagnosis

04



MultNAT® SYSTEM

Therapy-Oriented
Diagnosis



About Ustar




USTAR, founded in 2005, is a leading point-of-care molecular diagnosis provider for tuberculosis, COVID-19, malaria and other major communicable diseases.

Our vision and mission is:

Molecular Testing, Anywhere!



Milestone





Our Team



600+
Employees




over 50% have Master's Degrees and above

200+
Employees in R&D team





Good Manufacturing Practice



20,000+ m²


GMP workshop

In compliance with
ISO13485:2016





Global Presence



Distributors

500+

Installations

30,000+

Hospitals

3,000+

Countries

70+

International clients & partners

100+



Key Accounts



THE WORLD BANK
IBRD • IDA | WORLD BANK GROUP



CNPC



中國建築
CHINA STATE CONSTRUCTION



Abbott



HUAWEI



QUIDEL



Patents



Domestic and International Patents

90+

Granted Patents

45+






Key Products



PortNAT® System
At-Home Self Testing



EasyNAT® System
Fast Screening, Primary Diagnosis



MultNAT® System
Therapy-Oriented Diagnosis



CONTENTS

01



ABOUT USTAR

02



PortNAT® SYSTEM

At-Home
Self-Testing

03



EasyNAT® SYSTEM

Fast Screening and
Primary Diagnosis

04



MultNAT® SYSTEM

Therapy-Oriented
Diagnosis

PortNAT[®] System



Nucleic Acid Self-Testing



Result in 15 min



Simplified testing steps



Accurate


Nucleic acid Testing



Environmentally-friendly


Independent test cassette &
reusable incubator

PortNAT[®] System



Combine the speed of antigen test and the accuracy of PCR test

Workflow



Collect sample


Add sample

Amplification

Add buffer

Read result

Global Presence



Exported to **30+**



**countries and
regions, including**

Germany, Netherlands,
Italy, Spain,
Sweden, Israel,
HongKong, Japan,
Australia, Singapore,
Malaysia,
Brazil, Peru.




Application Scenarios



Business



Healthcare



Public Sector



Pet Clinic



Education



Airports & Customs



Testing Lab



Home Detection



Extensive Testing



- COVID-19
- MTC/NTM
- Flu A/B
- RSV
- LP
- Parainfluenza virus



- *Helicobacter pylori*
- *Feline herpes virus*
- *Feline calicivirus*
- *Feline mycoplasma*



- HIV
- TP
- CT/NG
- GBS
- Mycoplasma /Chlamydia



- *Salmonella*
- *L. monocytogenes*
- *S.aureus*
- STEC
- *Vibrio cholerae*



COVID-19 RNA Test

All-In-One Testing

Compact&Portable

Easy to Use

Test Anywhere,
Anytime



**COVID-19
RNA Test**

LOD: 3,000 copies/mL

Result in 35 min

PCR Quality

Variants Detected

Alpha/Beta/Delta/Omicron variants

COVID-19 RNA Test

First Nucleic Acid Self Test for COVID-19

Approved by:

- EU (CE)
- Australia (TGA)



Clinical Performance-Europe

Comparing with RT-PCR : Sensitivity \geq 95.4%
Specificity \geq 99.8%

Sensitivity	95.413%
Specificity	99.803%
Positive Likelihood Ratio	484.697
Negative Likelihood Ratio	0.046
Positive Predictive Value	99.048%
Negative Predictive Value	99.023%
Accuracy	99.028%

Study timeline: 2021.9-2022.1
 General Hospital Jesenice (Slovenia)
 Clinical Hospital Rijeka (Croatia)

Report No. 2022/01

15. Confusion Matrix

Reagents to be evaluated	PCR (Nasopharyngeal Sample)		Total
	Positive	Negative	
Positive	104	1	105
Negative	5	507	512
Total	109	508	617

Sensitivity	95.413%	89.619% to 98.494%
Specificity	99.803%	98.908% to 99.995%
AUC	0.969	0.961 to 0.987
Positive Likelihood Ratio	484.697	68.376 to 3435.875
Negative Likelihood Ratio	0.046	0.020 to 0.108
Positive Predictive Value	99.048%	93.619% to 99.865%
Negative Predictive Value	99.023%	97.731% to 99.583%
Accuracy	99.028%	97.895% to 99.642%

CT Range	CT \leq 25	25 < CT \leq 28	28 < CT \leq 31	CT > 31
No. of Cases (RT PCR)	30	14	28	37
No. of Positive Candidate Test Results	30	12	28	34
Positive Coincidence Rate	30/30	12/14	28/28	34/37

16. Incidents, Protocol Deviations, Excluded Samples and Rejected Results

No incidents occurred or were documented during the study testing.

There were no protocol deviations.

During the study one sample was excluded.

During the study no results were rejected.

Version 1.0
 SPLOŠNA BOLNIŠNICA JESENICE
 ENOTA ZA MIKROBIOLOGIJU
 Cesta maršala Tita 112, 4270 JESENICE
 Tel.: 04/5868218, fax: 04/5868496
 mat. št. 5053692, Id. za DDV/SI86143824

Miha Skvarč, dr. med.
 spec. klinične mikrobiologije

Page: 25

Clinical Performance-China

Tested by national reference:

Results of positive, negative, sensitivity, and repeatability reference panels all **met the requirements.**



中国食品药品检定研究院

检验报告

报告编号: RH202202873

检品名称: 新型冠状病毒(2019-nCoV)核酸检测试剂盒(自驱动微流控恒温扩增-试纸条法)

生产单位/产地: 杭州优思达生物技术有限公司

检验目的: 合同检验

检验依据: 产品技术要求



中国食品药品检定研究院检验报告

报告编号: RH202202874 共2页, 第2页

检验项目	标准规定	检验结果
	S1~S3应为阳性; S4~S10不作要求	S1~S5为阳性
	以下空白	

备注: 合同检验是在双方自愿基础上, 按照合同约定开展的样品检验。1、检验用参考品为新型冠状病毒核酸检测试剂国家参考品, 批号370099-202001, 由中国食品药品检定研究院提供。2、根据企业出具的说明, 本试剂盒未设置针对人源基因的内标检测体系; 检测靶基因为ORF1ab, 因此P7检测结果应为阴性。3、根据试剂盒“检验结果的解释”, 1) 结果阅读窗口“Test”区域出现清晰可见的红色线条, 则表示样本中检测到新型冠状病毒核酸, 即为阳性。在阳性结果中, 有时“Control”区域的线条较淡, 结果仍然有效; 2) 结果阅读窗口中只有“Control”区域出现红色线条, “Test”区域没有可见的红色线条, 则样本中未检测到新型冠状病毒核酸, 即为阴性; 3) 结果阅读窗口的两个区域中都没有可见的红色线条, 则测试失败, 其结果无效, 需重新检测。4、根据产品技术要求和企业出具的说明, 试剂盒无需进行病毒RNA提取纯化, 直接对样本进行检测; 在检测装置加样孔中加入国家参考品样本30 μL (P7样本加样量为30 μL), 打开一管缓冲液A液泡, 全部加入加样孔中, 然后盖上孔盖, 进行检测。本试剂盒为肉眼判读结果, 因此产品技术要求中未对精密性的变异系数进行要求。

检验结论	本品按产品技术要求检验, 结果符合规定。		
授权签字人		签发日期	2022年1月21日

Study unit: National Institutes for Food and Drug Control

Clinical Performance-China

1. Clinical trial in Disease Control and Prevention Center, Hubei, China:
 - ✓ **98.9% agreement rate with PCR assay (Ct < 35)**
 - ✓ Four types of SARS-CoV-2 strains (wild type, Alpha, Beta, Delta) were detected successfully
2. Clinical trial in State Key Laboratory for Diagnosis and Treatment of Infectious Diseases, The First Affiliated Hospital, Zhejiang University:
 - ✓ **98.5% agreement rate with comparison assay (ORF1ab Ct ≤ 35)**
 - ✓ Two SARS-CoV-2 strains from different sources were detected successfully.

Table 1: Ct Values for Various Samples

Sample ID	ORF1ab Ct	ORF2ab Ct	ORF3ab Ct	ORF4ab Ct	ORF5ab Ct	ORF6ab Ct	ORF7ab Ct	ORF8ab Ct	ORF9ab Ct	ORF10ab Ct
Sample 1	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 2	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 3	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 4	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 5	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 6	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 7	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 8	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 9	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Sample 10	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1

Table 2: Comparison of Detection Results

Strain	ORF1ab Ct	ORF2ab Ct	ORF3ab Ct	ORF4ab Ct	ORF5ab Ct	ORF6ab Ct	ORF7ab Ct	ORF8ab Ct	ORF9ab Ct	ORF10ab Ct
Beta	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1
Alpha	28.5	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1	29.1



CONTENTS

01



ABOUT USTAR

02



PortNAT® SYSTEM

At-Home
Self-Testing

03



EasyNAT® SYSTEM

Fast Screening and
Primary Diagnosis

04



MultNAT® SYSTEM

Therapy-Oriented
Diagnosis



EasyNAT[®] System

4/8/16



EasyNAT[®]
**Nucleic Acid Amplification
and Detection Analyzer**



Highlights



Fast



Accurate



Simple



High
Throughput




Cost
effective



Safe




Mechanism & Workflow



Swab in, Result out

DO NOT NEED
VTM & BSC


- Viral transport medium (VTM)
- Biological safety cabinet (BSC)



2



3



Simplified Workflow Collect and handle sample in **ONE** step

Biological Safty Nucleic acid lysis, amplification, purification and detection in **ONE** enclosed cartridge

Application Scenarios



**Hospitals &
Clinics**



CDC



Customs



**Mobile
testing**



**Pet
clinic**



**Testing
Lab**

Extensive Testing

- **SARS-CoV-2**
- **MTC**
- **MTC/NTM**
- **Influenza A/B Virus**
- **MP**
- **RSV**
- **BP**
- **LP**

Respiratory
Tract



- **CT/NG**
- **MG**
- **UU**
- **HSV 1/2**
- **GBS**
- **Syphilis**
- **TV**
- **HPV**
- **Monkeypox**

Reproductive
Health



- **FCV**
- **FHV**
- **CF**
- **TOX**
- **CPV**
- **CDV**
- **CHV**

Pet



- **ASF**
- **TGEV**
- **M. bovis**
- **Rinderpest virus**

Vet



- **Alicyclobacillus acidoterrestris**
- **Staphylococcus aureus**
- **Salmonella**

Food
Safety



- **BX**
- **HLB**

Plant



EasyNAT[®] Certificate & Registration

China NMPA**European Union CE****U.S. FDA****Australia TGA****Brazil ANVISA****Singapore HSA****Philippines FDA****Malaysia MDA****Nigeria NAFDAC****Thailand****Peru****Israel****U.S. FOOD & DRUG
ADMINISTRATION****Establishment Registration Device Listing**[New Search](#)[Back To Search Results](#)

Proprietary Name:	Nucleic Acid Amplification and Detection Analyzer
Classification Name:	REAL TIME NUCLEIC ACID AMPLIFICATION SYSTEM
Product Code:	<u>00I</u>
Device Class:	2
Regulation Number:	<u>862.2570</u>
Medical Specialty:	Clinical Chemistry
Registered Establishment Name:	<u>USTAR BIOTECHNOLOGIES (HANGZHOU) LTD.</u>
Registered Establishment Number:	3009312616
Owner/Operator:	<u>Ustar Biotechnologies (Hangzhou) Ltd.</u>
Owner/Operator Number:	10077585
Establishment Operations:	Manufacturer

EasyNAT[®] COVID-19

- **Targets:** ORF1ab, N gene
- **Sample type:** swab, sputum
- **LOD** ≤ 200 copies/mL
- **Sensitivity** ≥ 98.0%, **specificity** ≥ 99.6%
- **All in one cartridge:** sample-in, result-out, within 20 min
- **Ambient transportation:** -25~30°C



EasyNAT® Clinical Evaluation

1

Israel MOH

Comparable with Thermo
kit and GeneXpert



2

Spain MOH Lab


100% sensitivity and
specificity



3

US Cabell Huntington Hospital

Accuracy - 100%



4


UK NHS

Sensitivity - 100 %
Specificity - 100 %





EasyNAT[®] COVID-19 in Healthcare Units




Ustar EasyNAT[®] Covid-19 applied in healthcare units at all levels

EasyNAT[®] COVID-19 in Mobile Detection Vehicle




EasyNAT[®] MTC



- **Target:** IS6110
- **Sample type:** sputum, BALF, urine, paraffin-embedded tissue sample, pus, CSF, gastric aspirate, ascites with hydrothorax
- **LOD** ≤100 CFU/mL
- **Sensitivity** ≥ 96.6%, **specificity** ≥ 98.6%
- **Integrated testing:** sample-in, result-out, result in 65 min
- **Ambient transportation:** -25~30°C

EasyNAT[®] MTC: Workflow


Sputum




① Collect in a sputum container

② Liquefy without opening lid


Paraffin section



① Add lysis buffer ② Autoclave 10 min



③ Add into cartridge



④ Detection



⑤ Report result

EasyNAT[®] MTC: Highlights



Fast testing

Optimized sample pre-treatment, shortened turnaround time (TAT)



Point-of-care testing

Applicable for tent hospital, airport, and mobile van (lab)



Test anytime

Independent modules, providing instant testing for every single sample



Biologically safe

All-in-one sealed cartridge

Self-contained sputum container

Free-of-xylene paraffin section pretreatment process

Clinical Evaluation: Sputum

“EasyNAT could be used as an initial test for PTB diagnosis due to its simplicity, rapid turnaround time, high sensitivity, and low cost.”

Of the 169 PTB patients


Methodology	EasyNAT	Smear	Culture	Xpert
Sensitivity	72.19%	32.54%	53.85%	61.54%

Of the 91 patients with culture-positive outcomes


Methodology	EasyNAT	Smear	Xpert
Sensitivity	93.40%	56.04%	91.21%

Emerg Microbes Infect. 2021; 10(1): 1530–1535. Published online 2021 Aug

1. doi: [10.1080/22221751.2021.1959271](https://doi.org/10.1080/22221751.2021.1959271)



Clinical Evaluation: Children Gastric Aspirate



	Sensitivity, % (n/N)		Specificity, % (n/N)	
	EasyNAT	Xpert	EasyNAT	Xpert
All enrolled children	22.6 (31/137)	26.3 (36/137)	98.0 (100/102)	99.0 (101/102)
Bacteriologically confirmed	60.0 (9/15)	60.0 (9/15)		
Probable TB	18.0 (22/122)	22.1 (27/122)		

“EasyNAT may therefore be useful as a suitable alternative method of childhood TB diagnosis based on its cost-effectiveness, speed, and accuracy.”

Clinical Evaluation: Pathological Tissue

Clinical site: Beijing Chest Hospital, Capital Medical University

Sample enrolled: 103

Compare with RT-PCR:

Positive agreement rate: 97.01%

Negative agreement rate: 97%

Overall agreement rate: 97.09%

	RT-PCR			
		Positive (CT≤36)	Negative	Total
EasyNAT® MTC	Positive	65	1	66
	Negative	2	35	37
	Total	67	36	103

EasyNAT[®] MTC/NTM



- **Targets:** IS6110 (MTC), 16s rDNA (NTM)
- **Sample type:** sputum, BALF, urine, paraffin-embedded tissue sample, pus, CSF, gastric aspirate, ascites with hydrothorax
- **LOD** ≤100 CFU/mL
- **Integrated testing:** sample-in, result-out, result in 65 min
- **NTM:**
 - *M. Kansas*, *M. Avium*, *M. intracellular*, *M. chelonae*
 - *M. Abscess*, *M. Fortuitans*, *M. Smegmatis*, *M. Marseillae*
 - *M. scrofula*, *M. Gordon*, *M. toad*, *M. ulcerate*
 - *M. mariae*, *M. haemophilus*, *M. apes*, *M. gastroticus*

EasyNAT[®] CT/NG

”



- **Safe:** CT+NG detection in an enclosed cartridge
- **Sample type:** urethral swab, endocervical swab
- **LOD** ≤ 400 copies/mL
- **Integrated testing:** sample-in, result-out, result in 30 min
- **Ambient transportation:** -25~30°C

EasyNAT[®] CT/NG Clinical Evaluation

Clinical sites (856 included samples):

- Zhejiang Provincial People's Hospital
- Dermatology Hospital of Southern Medical University
- Beijing Ditan Hospital, Capital Medical University

	Positive Agreement (%)	Negative Agreement (%)	Total Agreement(%)	Kappa
CT	99.57% 95% CI (97.61%, 99.92%)	99.36% 95% CI (98.36%, 99.75%)	99.42% 95% CI (98.64%, 99.75%)	0.985 95% CI (0.973-0.998) , P < 0.001
NG	96.20% 95% CI (91.96%, 98.25%)	99.71% 95% CI (98.96%, 99.92%)	99.07% 95% CI (98.17%, 99.53%)	0.969 95% CI (0.947-0.990) , P < 0.001

EasyNAT[®] MG Clinical Evaluation

Clinical sites (787 included samples)

- Dermatology Hospital of Southern Medical University
- Dermatology Hospital of Jiangxi Hospital
- Tianjin Academy of Traditional Chinese Medicine Affiliated Hospital

Compared with PCR kit

		PCR kit		Total
		Positive	Negative	
EasyNAT [®]	Positive	156	0	156
	Negative	3	628	631
Total		159	628	787

Positive percent agreement=98.11%, 95% CI (94.58%~99.61%);

Negative percent agreement=100%, 95% CI (99.41%~100%);

Overall percent agreement=99.62%, 95% CI (98.89%~99.92%);

Kappa=0.9881, 95% CI (0.9746~1.0015); P < 0.001

EasyNAT[®] MP Clinical Evaluation

Clinical sites (665 included samples)

- The Affiliated Hospital of Hangzhou Normal University
- Children's Hospital of Hebei Province
- The Second Xiangya Hospital of Central South Hospital
- Jinan Children's Hospital

Compared with PCR kit

		PCR kit		Total
		Positive	Negative	
EasyNAT [®]	Positive	223	2	225
	Negative	2	438	440
	Total	225	440	665

Positive percent agreement=99.11%, 95% CI (96.83%~99.89%);

Negative percent agreement=99.55%, 95% CI (98.37%~99.94%);

Overall percent agreement=99.40%, 95% CI (98.47%~99.84%);

Kappa=0.9866, 95% CI (0.9734~0.9997); P < 0.001

EasyNAT® Flu A/B, B19, RSV Performance Evaluation

检 验 报 告

报告编号: W-W-0613-2021

委托方 杭州优思达生物技术有限公司

样品名称 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)

型 号 /

检验类别 注册检验 ()

其他检验 () 委托检验 ()

国家食品药品监督管理局北京医疗器械质量监督检验中心

EasyNAT Human Parvovirus B19 Assay

Test by: National Medical Products Administration,
Beijing Institute of Medical Device Testing

国家食品药品监督管理局北京医疗器械质量监督检验中心
检 验 报 告 首 页

报告编号: W-W-0613-2021

产品名称 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)

规格 / 型号规格 48 测试盒

委托方 杭州优思达生物技术有限公司

受托方 杭州优思达生物技术有限公司

生产单位 杭州优思达生物技术有限公司

检验单位 浙江省药品监督管理局

检验地点 杭州优思达生物技术有限公司

检验日期 2021 年 04 月 27 日

检验日期 2021 年 04 月 27 日

检验项目 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法) 医疗器械产品技术要求

检验结果 经样品检测, 符合《人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)》产品技术要求。

备注 1. 报告中的“—”表示该指标不适用, 报告中“*”表示该指标不适用。
2. 该产品技术要求中部分指标与国家注册标准、行业标准、行业标准不一致, 以该产品技术要求为准。
3. 2.2、2.3、2.4、2.5 项中国食品药品检定研究院人细小病毒 B19 核酸检测试剂盒国家注册标准 (型号: JY005-201601)。

检验员 王会会 审核 李海 批准 李海

中国食品药品检定研究院

MA 180000100599

中国食品药品检定研究院

检验报告

报告编号: RZ202103166

样品名称 人细小病毒 (B19) 核酸检测试剂盒 (恒温扩增-实时荧光法)

生产单位/产地 杭州优思达生物技术有限公司

检验目的 注册检验 (新产品首次注册/首次注册/质量注册复验)

检验依据 产品技术要求

EasyNAT Rapid RSV Assay

Test by National Institutes for Food and Drug Control

中国食品药品检定研究院检验报告

报告编号: RZ202103168

检验项目 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)

规格 / 型号规格 48 测试盒

委托方 杭州优思达生物技术有限公司

受托方 杭州优思达生物技术有限公司

生产单位 杭州优思达生物技术有限公司

检验单位 中国食品药品检定研究院

检验日期 2021 年 04 月 27 日

检验日期 2021 年 04 月 27 日

检验项目 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法) 医疗器械产品技术要求

检验结果 经样品检测, 符合《人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)》产品技术要求。

备注 1. 报告中的“—”表示该指标不适用, 报告中“*”表示该指标不适用。
2. 该产品技术要求中部分指标与国家注册标准、行业标准、行业标准不一致, 以该产品技术要求为准。
3. 2.2、2.3、2.4、2.5 项中国食品药品检定研究院人细小病毒 B19 核酸检测试剂盒国家注册标准 (型号: JY005-201601)。

检验员 王会会 审核 李海 批准 李海

中国食品药品检定研究院

MA 180000100599

中国食品药品检定研究院

检验报告

报告编号: RZ202103167

样品名称 人细小病毒 (B19) 核酸检测试剂盒 (恒温扩增-实时荧光法)

生产单位/产地 杭州优思达生物技术有限公司

检验目的 注册检验 (新产品首次注册/首次注册/质量注册复验)

检验依据 产品技术要求

EasyNAT Rapid Flu Assay

Test by National Institutes for Food and Drug Control

中国食品药品检定研究院检验报告

报告编号: RZ202103169

检验项目 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)

规格 / 型号规格 48 测试盒

委托方 杭州优思达生物技术有限公司

受托方 杭州优思达生物技术有限公司

生产单位 杭州优思达生物技术有限公司

检验单位 中国食品药品检定研究院

检验日期 2021 年 04 月 27 日

检验日期 2021 年 04 月 27 日

检验项目 人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法) 医疗器械产品技术要求

检验结果 经样品检测, 符合《人细小病毒 (B19) 核酸检测试剂盒 (PCR-荧光探针法)》产品技术要求。

备注 1. 报告中的“—”表示该指标不适用, 报告中“*”表示该指标不适用。
2. 该产品技术要求中部分指标与国家注册标准、行业标准、行业标准不一致, 以该产品技术要求为准。
3. 2.2、2.3、2.4、2.5 项中国食品药品检定研究院人细小病毒 B19 核酸检测试剂盒国家注册标准 (型号: JY005-201601)。

检验员 王会会 审核 李海 批准 李海

Thenational reference materials were used for registration and testing, and the results of positive, negative and sensitivity reference panels **all met the requirements.**



CONTENTS

01



ABOUT USTAR

02



PortNAT® SYSTEM

At-Home
Self-Testing

03



EasyNAT® SYSTEM

Fast Screening and
Primary Diagnosis

04



MultNAT® SYSTEM

Therapy-Oriented
Diagnosis

MultNAT[®] System

MultNAT[®] MOLECULAR DIAGNOSTIC PCR SYSTEM



FDA-Listed



Multiplex



High-throughput



Bio-safe



Test to Treat

Diagnosis provides guidance on medication and therapy

- ① Diagnosis of infection of unknown origin (Respiratory, reproductive tract, Gastrointestinal, CNS, Blood)
- ② Individual medication differences (Warfarin , Clopidogrel)
- ③ Quantitative testing (HIV, HBV)
- ④ Antimicrobial resistance (MTC, CRE)



MultNAT[®] System: Highlights

※ Extensive panels

Test for virus, bacteria, parasites, genotyping, and antimicrobial resistance

※ Multiplex

1 sample, up to 24 targets per round

※ High-throughput

1 test, up to 32 samples per round

※ Safe

All in one enclosed cartridge

※ User-friendly

13.3 inches touch screen & QR code access


MultNAT[®] Certificate & Registration

FDA U.S. FOOD & DRUG ADMINISTRATION | Establishment Registration Device Listing

Proprietary Name:	MultNAT Molecular Diagnostic Testing System; Nucleic Acid Amplification and Detection Analyzer
Classification Name:	REAL TIME NUCLEIC ACID AMPLIFICATION SYSTEM
Product Code:	<u>OOI</u>
Device Class:	2
Regulation Number:	<u>862.2570</u>
Medical Specialty:	Clinical Chemistry
Registered Establishment Name:	<u>USTAR BIOTECHNOLOGIES (HANGZHOU) LTD.</u>
Registered Establishment Number:	3009312616
Owner/Operator:	<u>Ustar Biotechnologies (Hangzhou) Ltd.</u>
Owner/Operator Number:	10077585
Establishment Operations:	Manufacturer



MultNAT[®] System: Consumables



- 1 Sample in, result out
- 2 High specificity
- 3 Extract, purify, amplify and detect nucleic acid in a sealed tube
- 4 Pre-filled reagent
- 5 Glassified reagent enables ambient transport
- 6 No cross contamination

MultNAT[®] System: Consumables



1-Strip Tube




8-Strip Tubes



2-Interconnected
tube




4-Interconnected
Tube



8-Interconnected
Tube




MultNAT[®] System: Workflow



MultNAT[®] System: Test Menu

Pathogen Infection

- RTI panel
- STI panel
- Meningitis panel
- Gastrointestinal panel
- BCID panel



Medication Guidance

- MTC/MDR
- MTC/XDR
- Carba-R
- MRSA/SA
- Warfarin Gene Polymorphism
- *H. pylori* resistance

Genotyping

- NTM
- HPV
- HIV

Viral Load Testing

- *Epstein-Barr virus*
- HIV
- HBV
- HCV

MultNAT® Carba-R and Four Pathogens PCR Assay (RUO)

Mod	Target	Fluorescent channel	Sensitivity			
A	OXA	FAM	OXA-48 (<i>K.pneumoniae</i>)	100 CFU/mL	OXA-48 (<i>E.Coli</i>)	50CFU/mL
	KPC	HEX	KPC-3 (<i>K.pneumoniae</i>)	50CFU/mL	KPC-4 (<i>K.pneumoniae</i>)	100CFU/mL
	VIM	ROX	VIM-4 (<i>E.Coli</i>)	50CFU/mL	VIM-1 (<i>K.pneumoniae</i>)	50CFU/mL
B	IMP	FAM	IMP-1 (<i>E.Coli</i>)	50CFU/mL	IMP-2 (<i>P.aeruginosa</i>)	100CFU/mL
C	NDM	FAM	NDM-1 (<i>K.pneumoniae</i>)	200CFU/mL	NDM-2 (<i>A.baumannii</i>)	50 CFU/mL
	<i>K.pneumoniae</i>	HEX	400 CFU/mL		/	
	<i>A.baumannii</i>	ROX	400 CFU/mL		/	
D	<i>P.aeruginosa</i>	FAM	400 CFU/mL		/	
	<i>E.Coli</i>	ROX	400 CFU/mL		/	

MultNAT® Carba-R and Four Pathogens PCR Assay (RUO)

Cmpare with
PCR (Xpert)

Target	Template	Test concentration CFU/mL	MultNAT/Ct	Xpert/Ct
OXA	OXA-48 (<i>K.pneumoniae</i>)	1	N/A	N/A
		10	N/A	N/A
		100	37.7	37.6
IMP	IMP-1 (<i>E.Coli</i>)	10	N/A	N/A
		100	33.9	39.0
		1000	30.8	33.3
KPC	KPC-4 (<i>K.pneumoniae</i>)	10	N/A	N/A
		100	35.9	36.7
		1000	30.8	29.6
VIM	VIM-1 (<i>K.pneumoniae</i>)	10	36.0	N/A
		100	36.5	37.6
		1000	31.9	30.4
NDM	NDM-2 (<i>A.baumannii</i>)	10	N/A	N/A
		100	36.0	37.0
		1000	34.6	37.7

MultNAT® Gastrointestinal Panel (RUO)

Mod	Pathogens	Fluorescent channel	Sensitivity (cps/mL)
A	Salmonella	HEX	100
	Campylobacter	CY5	1000
	<i>E. Coli</i> O157	FAM	1000
B	<i>Y. enterocolitica</i>	HEX	30
	<i>C. difficile</i> (Toxins A/B)	CY5	1000
	<i>Enterotoxigenic E. coli</i> (ETEC)	FAM	5000
C	Shiga Toxin-Producing <i>Escherichia coli</i> (STEC)	HEX	300
	<i>Shigella Castellani</i> (plus EIEC)	CY5	100
	<i>Vibrio</i> (<i>V. cholerae</i> , <i>V. parahaemolyticus</i> , <i>V. vulnificus</i>)	FAM	100
D (Virus)	Adenovirus 40/41	HEX	1000
	<i>Rotavirus</i>	CY5	100
	<i>Norovirus</i> GI/GII	FAM	1000

Internal control is at β -actin, ROX channel

Highlights

- **12 pathogens tested in ONE cartridge**
- **Less interference**
- **Higher specificity & sensitivity**

A group of six young boys are sitting on a low wall made of red bricks. They are all smiling and looking towards the camera. The boy on the far left is wearing glasses and a pink shirt. The boy next to him is wearing a dark blue shirt. The boy in the middle is wearing a light-colored shirt. The boy next to him is wearing an orange shirt. The boy next to him is wearing a yellow shirt. The boy on the far right is wearing a blue denim shirt. They are all wearing blue pants. The background shows a light blue wall, a tree, and some outdoor furniture.

MOLECULAR TESTING ANYWHERE

THANKS



6/F, Building 2, 611 Dongguan Road, Binjiang
District, Hangzhou, Zhejiang, 310053 PR China



+86-571-88939358



en.bioustar.com



sales@bioustar.com

